


Getinge Assured

Bowie-Dick Test Card TT

The Getinge Assured Bowie-Dick Test Card TT is for daily monitoring of small pre-vacuum type B steam sterilizers operating at 134°C. The card is specially packaged and designed to meet EN 867-5 requirements.

This Getinge Assured product is tested and verified to have equivalent and approved performance at both EN 867-5 test cycles and Getinge Bowie-Dick cycles for type B pre-vacuum table top sterilizers.



Performance Characteristics

If the sterilizer is functioning properly, the indicator graphic will show a uniform color change from purple to green throughout the entire indicator image. A brown or non-uniform green color may indicate problems with steam conditions in the sterilizer chamber other than non-condensable gases or residual air. This may require a sterilizer service call.

Physical Characteristics

• mensions: L 86mm x W 86mm

Packaging: 1 ring and 15 test cards

Environmental Limitations

Storage: In provided sealed packaging at room temperature 10°-38°C and 10-60% Relative Humidity. Protect from moisture and excess humidity.

Shelf Life

24 months. The expiry date is printed on the product packaging.

Standards Classification

EN 867-5

Ordering Information

Getinge Assured

Bowie-Dick Test Card TT

Order Number

504052300

PRS528_02_EN - Getinge Assured Bowie-Dick Test Card TT

Getinge Infection Control AB, P O Box 69, SE-305 05 Getinge, Sweden www.getinge.com



Team Medical SuppliesPhone 1300 22 44 50

Fax 1300 22 44 60 info@teammed.com.au www.teammed.com.au



BP Medical (WA)

Phone 1300 33 55 60 Fax 1300 33 55 70 info@bpmedical.com.au www.bpmedical.com.au