


Instructions For Use

Mid Range 1L / 1.8L / 2L / 4L / 5L / 7L

1

Assemble the lid and base with fill line and label to the front


This ensures you can always see when the container is full.



2

Click down all four corners

Assemble on a flat stable surface and FIRMLY push down each corner until a click is heard.



3

Fill in the label before and after use


This ensures a permanent audit trail.



4

Never fill above the line

Always replace your sharps container when full to avoid needle stick injuries.




5

Temporary closure

The temporary closure feature should always be engaged when the container is not in use. To engage, close the lid until it covers the aperture and a click is heard.

To release, pull back on the tab and lift the lid until it engages in the open position.



6

Final closure

To engage final closure, place the container on a stable surface.

With the container in temporary closure, press the lid down firmly ensuring all three clicks are heard and tab sits within the recess.

Container is now ready for disposal.



FOR MORE INFORMATION:

✉ sales@plascare.com.au

🌐 sharpsafe.com.au 📞 +61 3 9458 5118